

YZ38 Keratometer



The product is applied in measuring diopter on the surface of cornea, and the axis and diopter of cornea astigmatism. It can measure the diopter of implanting lens assisted by A ultrasonic.

Specifications:

Measuring Range	Curvature Radius	5.511mm
	Diopter	30~60D
Minimum Division	Radius	0.02mm
	Diopter	0.25D
Minimum Surface Area Required by Measurement	r = 5.5mm	1.65mm
	r = 7.5mm	2.36mm
	r = 11mm	3.36mm